



AICRP ON AGROFORESTRY
COLLEGE OF FORESTRY, DAPOLI
Dr. Balasaheb Sawant Konkan Krishi Vidyapeeth, Dapoli. Tal. Dapoli,
Dist: Ratnagiri - 02358 240340



No.CFOR/AF/ / ~~877~~ /of 2019

AICRP on Agroforestry
 College of Forestry, Dapoli

Dated: **31 OCT 2019**

To,

Subject: - Extension of date for the quotation of supply of UV- spectrophotometer

Sir,

It was decided to purchase the following equipment vide this office quotation letter No. CFOR/AF/752/of 2019 dated 03 Oct 2019. However, this office has not received any quotation upto prescribed date of 30/10/2019. It is therefore, decided to extend the last date of receipt of quotation i.e. upto 20/11/2019. It is therefore, requested to kindly quote your comparative rates for supply of following item. The detail technical specification is as follows.

Sr.No.	Particulars	Approximate Quantity required
1	<p>UV- Spectrophotometer (SPECTRONIC 200, 100-250V, US Power Cord) Includes:</p> <ul style="list-style-type: none"> ✓ Full on board software with analyzer mode ✓ Readout precision of 0.000A to 0.999A and 1.00A to 2.50A ✓ Two cuvette racks ✓ 12 plastic cuvettes <p>Removable power cord; Rubber feet on the back;Store on a bookshelf. Spectral Bandwidth: <= 4 nm Light Source: Tungsten-halogen Detector: 2048 element CCD WavelengthRange: 340 nm –1000 nm Accuracy: ± 2 nm Repeatability: ± 1 nm Data interval: 1 nm Photometric Range: -0.3 A to 2.5 A Readout: ABS, %T, Concentration Accuracy: ± 0.01 A at 0.3 A ± 0.05 A at 1.0 A (SPECTRONIC standard filters measured at 590 nm) Repeatability: ± 0.3 %T at 50 %T Stray Light : < 0.2 %T (with SPECTRONIC standard SRM 400filter) Display:Variable angle 320 x 240 pixel color graphical LCD 7 × 5 cm, 8.6 cm diagonal (2.75" × 2", 3.4"diagonal) Keypad: Rubber Membrane</p>	01

The term and conditions of quotation are as follows:

1. The quotation should reach to the undersigned on or before 20/11/2019.
2. The quotation should include all the taxes, incidental charges, packing forwarding charges, etc.