



Dr. Balasaheb Sawant Konkan Krishi
Vidyapeeth, Dapoli
AICRP ON IRRIGATION WATER MANAGEMENT
CENTRAL EXPERIMENT STATION, WAKAWALI



Dr. R.T. Thokal
Chief Scientist

E-mail: dapoli.csiwm@gmail.com

No. CS/IWM/Quotation/Equip./ 424 /21

Date: 1 MAR 2021

To,

Sub: Supply of quotations for **Infrared Thermometer and LUX meter**

Sir,

With respect to above subject, it is proposed to purchase the **Infrared Thermometer and LUX meter** for AICRP on Irrigation Water Management, CES, Wakawali. For the supply of this equipment, the quotations are invited in the sealed envelope to this office on or before _____. Quotations should be addressed to **Chief Scientist, AICRP-IWM, Central Experiment Station, Wakawali, Tal. Dapoli, Dist. Ratnagiri, MS 415 712**. Quotations received after due date will not be considered for further purchase process. The **rates of equipment should be inclusive of all taxes, charges and F.O.R. at CES, Wakawali**. The quotation calling letter is also available on University website www.dbskkv.org till the last date.

Please write on the envelope as "Infrared Thermometer and LUX meter"

Details of equipment

Sr. No.	Name of equipment	Specifications	Rate per unit (Rs.)
1	Infrared thermometer	Temperature range: -50°C-1050°C or -58°F-1922°F Accuracy: ±1.5°C Distance spot ratio: 50:1 Wavelength: 630-670 nm Spectral response: 8 µm -14 µm Resolution: 0.1°C/0.1°F LCD size: 36 x 27 mm Power: 9V battery Facilities: Auto power off, low battery indicator, backlight display, MIN function, DIF function, Data	


		storage. High/low temp. alarm setting function Accessories: Carry box, Instruction manual, Tripod stand, Installed battery (9V)	
	Digital LUX meter	Range: 0-200,000 lux Accuracy: $\pm 5\%$ Resolution: 0.1 lux Power: 9V battery Selection: lux/fc Accessories: Carrying case, Manual, 9V battery (installed)	

The quotations should be submitted by abiding the following terms and conditions

Terms and conditions

1. Quotations in sealed envelope should reach to the office of Chief Scientist, AICRP-IWM, CES, Wakawali within mentioned due date.
2. Quotations on Xeroxed letter head will not be accepted by the office
3. Quotations received by the office within Deadline will be opened on the next working day in working hours.
4. GST number of the Company/Firm, PAN/TIN Number and Bank account details should be clearly indicated/enclosed with quotation.
5. This office will not be responsible for any postal delay.
6. Brochures of equipment indicating technical details should be enclosed with quotations.
7. In case of Manufacturer, the Manufacturer's Certificate should be enclosed with quotations. In case of Distributor/Dealer, the Authorization Certificate by the Company should be attached with Quotation.
8. All the rights of rejection or modification or extension of deadline of quotation calling letter/process without any reason are reserved by this office.
9. Binding of Repair and maintenance of the equipment will be on supplier for at least **THREE** years.

Yours sincerely,


Chief Scientist,

AICRP-IWM, CES, Wakawali

Copy with request to upload on University website for wide publicity to
In Charge, Agricultural Knowledge Management Unit, DBSKKV, Dapoli