

## TARAPOREVALA MARINE BIOLOGICAL RESEARCH STATION

(Dr. B.S. Konkan Krishi Vidyapeeth, Dapoli)
3<sup>rd</sup> Floor, New Administrative Building, Bandra (E), Mumbai 400051.
Tel/Fax: (022) 26516816; E-mail. tmbrs@rediffmail.com

No. TMBRS/QUOTATION/ Kjeldahl Unit / ......................../2021. Dated: 0 4 OCT 2021

Ţo,	

Sub: Supply of quotations for Kjeldahl digestion and distillation unit. .

Sir/Madam,

This office is desirous of purchasing / fabricating and installing of a **Kjeldahl digestion and distillation unit.** The details of the same are attached herewith. You are requested to quote your competitive rates, item – wise for the purchase/supply and installation etc. in detail, as per the below mentioned terms and conditions:

## **Terms & Conditions:**

- Quotation should be sent in a closed Envelope super scribing "Quotation for the purchase of Kjeldahl digestion and distillation unit." by post or courier only. Hand delivery of the quotations will not be accepted.
- 2. The quotation should clearly state the Basic price, applicable taxes and other charges if any. It is necessary to mention **GST number** on quotations.
- 3. Only those items whose physical delivery can be effected within should be included. Quotation must bear written declaration to the effect.
- 4. Terms and conditions regarding payment, delivery and guarantee should be mentioned clearly. However, the delivery of the material should be sent to: Research Officer, Taraporevala Marine Biological Research Station, 3<sup>rd</sup> Floor, Administrative Building, Bandra (E), Mumbai 400051; and your payment will be made against the delivery. Please note that any Bank formalities will not be accepted.
- 5. The Demand Draft Commission as well as postage regarding Payment is to be borne by the supplier.
- 6. The delivery should be FOR Destination (as the Research Station/College is situated in the Municipal limits 'Octrai' is not applicable).
- 7. If you are the Sole Manufacturer or appointed as an Authorized Dealer for the sale, please enclose a copy of certificate stating as Sole Manufacturer / Authorized Dealer.
- 8. The undersigned reserves the right to reject / withdraw/correct the content/ or to extend the date of acceptance for any or all quotations without giving any reason thereof.
- 9. Quotation complete in above details should be sent on or before 20.10.2021 up to 5.00 PM to:

The Research Officer,

Taraporevala Marine Biological Research Station,

3rd Floor, New Administrative Building, Bandra (E), Mumbai, (MS) - 400 051.

Yours sincerely,

Research Officer Research Officer

Copy submitted to the:

- 1. The Director of Research, Dr.B.S. Konkan Krishi Vidyapeeth, Dapoli, Dist. Rathagir 51
- 2. The Director of Extension, Dr.B.S. Konkan Krishi Vidyapeeth, Dapoli, Dist. Ratnagiri.
- 3. The Manager, Agriculture Knowledge Management Unit (AKMU), DBSKKV, Dapoli.
  - You are requested to publish said "Quotation Calling" letter/advertisement on the University website: <a href="https://www.dbskkv.org">www.dbskkv.org</a>.
- 4. Notice Board, TMBRS, Mumbai

Detailed specifications of the - Kjeldahl digestion and distillation unit.

Sr. no.	Kjeldahl digestion and distillation unit.	Quantity required
1.	Salient Features	
	Electrically heated unit	
	<ul> <li>Comprising of single mild steel tubular stand for both digestion &amp; distillation parts</li> </ul>	
	<ul> <li>Stand to be equipped with 2 different heating shelves, one above the other.</li> </ul>	
	Complete with heating box and fume duct.	
	<ul> <li>Specially designed fume duct to totally avoid leakage of Fumes</li> </ul>	
	<ul> <li>B-24 cone adapters openings suitable to fit Kjeldhal flasks having B-24 Sockets</li> </ul>	
	<ul> <li>Should be equipped with 500 watt heaters - in both the heating shelves.</li> </ul>	
	<ul> <li>Individual energy regulator for each heater in both the heating shelves</li> </ul>	
	<ul> <li>Distillation part to be provided with brass condenser rack and water circulation systems</li> </ul>	
	<ul> <li>Both distillation part and digestion part to be supplied with suitable number if clamps</li> </ul>	
	<ul> <li>Electrical operation: on 230 Volt, 50 Hz, Single Phase, AC Supply</li> </ul>	

Principal Investigator
TMBRS, Mumbai